UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/670,618	09/25/2003	Stephen T. Flock	D6476	6784
Benjamin Aaro	7590 02/10/200 n Adler	EXAMINER		
ADLER & ASSOCIATES			WITCZAK, CATHERINE	
8011 Candle Lane Houston, TX 77071			ART UNIT	PAPER NUMBER
			3767	
			MAIL DATE	DELIVERY MODE
			02/10/2009	PAPER

Please find below and/or attached an Office communication concerning this application or proceeding.

The time period for reply, if any, is set in the attached communication.

	Application No.	Applicant(s)
	10/670,618	FLOCK ET AL.
Office Action Summary	Examiner	Art Unit
	CATHERINE N. WITCZAK	3767
The MAILING DATE of this communication ap Period for Reply	ppears on the cover sheet with the o	correspondence address
A SHORTENED STATUTORY PERIOD FOR REPLEWHICHEVER IS LONGER, FROM THE MAILING ID. - Extensions of time may be available under the provisions of 37 CFR 1 after SIX (6) MONTHS from the mailing date of this communication. - If NO period for reply is specified above, the maximum statutory period. - Failure to reply within the set or extended period for reply will, by stature Any reply received by the Office later than three months after the mailing earned patent term adjustment. See 37 CFR 1.704(b).	DATE OF THIS COMMUNICATION .136(a). In no event, however, may a reply be tired will apply and will expire SIX (6) MONTHS from the cause the application to become ABANDONE	N. nely filed the mailing date of this communication. ED (35 U.S.C. § 133).
Status		
Responsive to communication(s) filed on 16 c This action is FINAL . 2b) ☐ This action is FINAL . Since this application is in condition for allowed closed in accordance with the practice under	is action is non-final. ance except for formal matters, pro	
Disposition of Claims		
4)	28,43,44,50 and 51 is/are withdraw	
<u> </u>		
9) The specification is objected to by the Examin 10) The drawing(s) filed on is/are: a) ac Applicant may not request that any objection to the Replacement drawing sheet(s) including the correct 11) The oath or declaration is objected to by the E	cepted or b) objected to by the drawing(s) be held in abeyance. Se ction is required if the drawing(s) is ob	e 37 CFR 1.85(a). jected to. See 37 CFR 1.121(d).
Priority under 35 U.S.C. § 119		
12) Acknowledgment is made of a claim for foreig a) All b) Some * c) None of: 1. Certified copies of the priority documer 2. Certified copies of the priority documer 3. Copies of the certified copies of the priority documer application from the International Burea * See the attached detailed Office action for a lis	nts have been received. nts have been received in Applicat ority documents have been receive au (PCT Rule 17.2(a)).	ion No ed in this National Stage
Attachment(s) 1) Notice of References Cited (PTO-892) 2) Notice of Draftsperson's Patent Drawing Review (PTO-948) 3) Information Disclosure Statement(s) (PTO/SB/08) Paper No(s)/Mail Date	4) Interview Summary Paper No(s)/Mail D 5) Notice of Informal F 6) Other:	ate

Application/Control Number: 10/670,618 Page 2

Art Unit: 3767

DETAILED ACTION

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(e) the invention was described in (1) an application for patent, published under section 122(b), by another filed in the United States before the invention by the applicant for patent or (2) a patent granted on an application for patent by another filed in the United States before the invention by the applicant for patent, except that an international application filed under the treaty defined in section 351(a) shall have the effects for purposes of this subsection of an application filed in the United States only if the international application designated the United States and was published under Article 21(2) of such treaty in the English language.

1. Claim 1 is rejected under 35 U.S.C. 102(e) as being anticipated by Unger et al (US 2002/0099356).

Unger et al disclose a device comprising an actuator (paragraph 0066) for driving an abrasive surface (16), a container adapted to deliver an abrasive material (20); a reservoir adapted to contain a pharmaceutical (12); and a permeable member (22).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

- (a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negatived by the manner in which the invention was made.
- 2. Claims 34-37 are rejected under 35 U.S.C. 103(a) as being unpatentable over Unger et al as applied above and further in view of Avrhami et al (US 2003/0212397). Now even though Unger does not explicitly disclose monitoring feedback using an electrical property attention is directed to Avrhami et al.

Art Unit: 3767

The Avrhami et al reference teaches in paragraphs [0070], [0121], [0124-5], and [0173] monitoring feedback using a heartbeat to perform a safe ablation procedure. It would have been obvious to one of ordinary skill in the art at the time of the invention to use the teachings of Avrhami et al in the device of Unger to increase the safety of the ablation procedure for better patient outcome.

Response to Arguments

Applicant's arguments filed 1/16/2009 have been fully considered but they are not persuasive. Applicant argues that Unger et al fails to disclose a device having a container to collect ablated tissue or other bimolecule and thus does not anticipate the claim. However, Examiner points out that as written, the claim currently claims "a container operably connected distally to the device having an opening therethrough adapted to deliver an abrasive material therein to the tissue OR to collect ablated tissue or other bimolecule therein." It is the position that the Unger et al reference does disclose a container having an opening adapted to deliver an abrasive material therein to the tissue, and does not need to "collect ablated tissue or other biomolecule therein" since the claim makes it clear that the container can perform either one (not necessarily both) of these functions.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to CATHERINE N. WITCZAK whose telephone number is (571)272-7179. The examiner can normally be reached on Monday through Friday, 8-5 EST.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Kevin Sirmons can be reached on (571) 272-4965. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Application/Control Number: 10/670,618

Art Unit: 3767

Information regarding the status of an application may be obtained from the Patent Application

Page 4

Information Retrieval (PAIR) system. Status information for published applications may be obtained

from either Private PAIR or Public PAIR. Status information for unpublished applications is available

through Private PAIR only. For more information about the PAIR system, see http://pair-

direct.uspto.gov. Should you have questions on access to the Private PAIR system, contact the Electronic

Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer

Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR

CANADA) or 571-272-1000.

/Catherine N Witczak/

Examiner, Art Unit 3767

/Kevin C. Sirmons/

Supervisory Patent Examiner, Art Unit 3767